

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST M. GUILBAULT AGENCY GOODYEAR PD
DATE 8/12/2010 TIME 1545
INTOXILYZER SERIAL # 80-001858 LOCATION GPA #2

- (AP) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (AP) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
(AP) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
(AP) 4. Air blank completed.
(AP) 5. Calibration check completed. Test results 0. 101 AC.
(AP) 6. Air blank completed.
(AP) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE [Signature] 11411

INTOXILYZER 8000
Location: GOODYEAR PD
Serial Number: 80-001858
Core Version: 8105.48
08/11/2010

15:45:26

Standard Lot#: 526576
Last Changed By: M. GUILBAULT #1144

QAS: M. GUILBAULT #1144
GOODYEAR PD

Operator: M. GUILBAULT #1144
GOODYEAR PD

Test	g/210L	Time
Air Blank	0.000	15:47:09
0.100 Cal Check	0.101	15:47:29
Air Blank	0.000	15:47:59

Preventative Maintenance
Performed -- Timer Reset


1144